



## AssayDef Human C-Reactive Protein (CRP) Deficient Plasma (Sodium Citrate)

Catalog No. D100101  
Sample Insert/Reference Only

### Presentation

- 1 vial of Depleted Human Plasma (1 mL, Lyophilized)
- The plasma was established as >99% deficient on average by ELISA testing.
- Albumin level serves as control for depletion process

### Reconstitution

- Add 1 ml reagent grade water to produce 1 x deficient plasma

### Warnings

- The Deficient Plasma is for **Research Use Only**
- Variations between laboratories may be caused by technique differences. It is recommended that each laboratory establish its own acceptable ranges.
- Although all whole blood was deemed HIV and HBsAg negative, use universal precautions while testing.

### Storage Condition

- Store at <-20°C until the expiration date.
- After reconstitution divide into 0.5 or 1 mL aliquots and store at <-80°C until the expiration date.

