



## Human Complement 3 Deficient Plasma (Serum)

Catalog No. D210111  
Sample Insert/Reference Only

### Reagent Details

- 1 vial of Depleted Human Serum (Lyophilized)
- The plasma was established as 95% deficient on average by ELISA testing.
- Albumin level serves as control for depletion process

### Reconstitution

- Reagent is a 1x deficient plasma

### Warnings

- The Deficient Plasma is for **Research Use Only**
- Variations between laboratories may be caused by technique differences. It is recommended that each laboratory establish its own acceptable ranges.
- Although all whole blood was deemed HIV and HBsAg negative, use universal precautions while testing.

### Storage Condition

- Store at  $-60^{\circ}\text{C}$ , or below, until the expiration date.

