



AssayMax™
Human Aldose Reductase
ELISA Kit

Assaypro LLC
3400 Harry S Truman Blvd
St. Charles, MO 63301
T (636) 447-9175
F (636) 395-7419
www.assaypro.com

For any questions regarding troubleshooting or performing the assay, please contact our support team at support@assaypro.com.

Thank you for choosing Assaypro.

Assay Summary

Step 1. Add 50 μ l of Standard or Sample per well.
Incubate 2 hours.

Step 2. Wash, then add 50 μ l of Biotinylated Antibody per well.
Incubate 2 hours.

Step 3. Wash, then add 50 μ l of SP Conjugate per well.
Incubate 30 minutes.

Step 4. Wash, then add 50 μ l of Chromogen Substrate per well.
Incubate 20 minutes.

Step 5. Add 50 μ l of Stop Solution per well.
Read at 450 nm immediately.

Symbol Key



Consult instructions for use.

Assay Template

[illegible]

AssayMax™ Human Aldose Reductase (AR) ELISA Kit

Catalog No. EA2670-1

Sample insert for reference use only

Introduction

Aldose reductase (AR), also known as aldo-keto reductase family 1 member B1 or AKR1B1, is a member of the monomeric, NADPH-dependent aldo-keto reductase family. The gene codes for a 316-amino acid protein with a molecular mass of 36 kDa. It catalyzes the reduction of aldose to the corresponding sugar alcohol, in particular glucose to sorbitol, which is subsequently metabolized to fructose by sorbitol dehydrogenase. The conversion of glucose to fructose constitutes the polyol pathway of glucose metabolism. This pathway plays a minor role in glucose metabolism under normal physiological conditions. However, in hyperglycemia associated with diabetes, cells can produce significant quantities of sorbitol. The accumulation of sorbitol, by an increased flux through AR, is linked between hyperglycemia and the development of vascular and neurological complications of diabetes (1). AR regulates TNF-alpha-induced cell signaling and apoptosis in vascular endothelial cells (2).

Principle of the Assay

The AssayMax™ Human Aldose Reductase ELISA (Enzyme-Linked Immunosorbent Assay) Kit is designed for detection of aldose reductase in human **plasma and serum samples**. This assay employs a quantitative **sandwich enzyme immunoassay** technique that measures human aldose reductase in approximately 5 hours. A polyclonal antibody specific for human aldose reductase has been pre-coated onto a 96-well microplate with removable strips. Aldose reductase in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for human aldose reductase, which is recognized by a streptavidin-peroxidase (SP) conjugate. All unbound material is washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- This product is for **Research Use Only** and is not intended for use in diagnostic procedures.

- Prepare all reagents (diluent buffer, wash buffer, standard, biotinylated antibody, and SP conjugate), as instructed, prior to running the assay.
- Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this insert. However, the user should determine the optimal dilution factor.
- Spin down the SP conjugate vial, the biotinylated antibody vial, and the standard diluent vial before opening and using contents.
- The Stop Solution is an acidic solution.
- The kit should not be used beyond the expiration date.

Reagents

- **Human Aldose Reductase Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human aldose reductase.
- **Sealing Tapes:** Each kit contains 3 precut, pressure sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Human Aldose Reductase Standard:** Human aldose reductase in a buffered protein base (16 ng, lyophilized).
- **Biotinylated Human Aldose Reductase Antibody (40x):** A 40-fold concentrated biotinylated polyclonal antibody against human aldose reductase (150 µl).
- **EIA Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (20 ml).
- **Standard Diluent (1x):** A buffered protein base with stabilizer (2 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **SP Conjugate (100x):** A 100-fold concentrate (80 µl).
- **Chromogen Substrate (1x):** A stabilized peroxidase chromogen substrate tetramethylbenzidine (7 ml).
- **Stop Solution (1x):** A 0.5 N hydrochloric acid solution to stop the chromogen substrate reaction (11 ml).

Storage Condition

- Upon arrival, immediately store components of the kit at recommended temperatures up to the expiration date.
- Store Standard, SP Conjugate, and Biotinylated Antibody at -20°C.
- Store Microplate, Diluent Concentrate (10x), Standard Diluent (1x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C.
- Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 30 days in a vacuum desiccator.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes (1-20 μl , 20-200 μl , 200-1000 μl , and multiple channel)
- Deionized or distilled reagent grade water

Sample Collection, Preparation, and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x *g* for 10 minutes and collect plasma. The sample is suggested for use at 1x; however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles (EDTA or Heparin can also be used as an anticoagulant).
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 3000 x *g* for 10 minutes and remove serum. The sample is suggested for use at 1x; however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Applicable samples may also include biofluids, cell culture, and tissue homogenates. If necessary, user should determine optimal dilution factor depending on application needs.

Refer to Dilution Guidelines for further instruction.

Guidelines for Dilutions of 100-fold or Greater <i>(for reference only; please follow the insert for specific dilution suggested)</i>	
100x	10000x
A) 4 μl sample : 396 μl buffer (100x) = 100-fold dilution <i>Assuming the needed volume is less than or equal to 400 μl.</i>	A) 4 μl sample : 396 μl buffer (100x) B) 4 μl of A : 396 μl buffer (100x) = 10000-fold dilution <i>Assuming the needed volume is less than or equal to 400 μl.</i>
1000x	100000x
A) 4 μl sample : 396 μl buffer (100x) B) 24 μl of A : 216 μl buffer (10x) = 1000-fold dilution <i>Assuming the needed volume is less than or equal to 240 μl.</i>	A) 4 μl sample : 396 μl buffer (100x) B) 4 μl of A : 396 μl buffer (100x) C) 24 μl of B : 216 μl buffer (10x) = 100000-fold dilution <i>Assuming the needed volume is less than or equal to 240 μl.</i>

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- EIA Diluent Concentrate (10x):** Dilute the EIA Diluent Concentrate 10-fold with reagent grade water to produce a 1x solution. When diluting the concentrate, make sure to rinse the bottle thoroughly to extract any precipitates left in the bottle. Mix the 1x solution gently until the crystals have completely dissolved. Store for up to 30 days at 2-8°C.
- Human Aldose Reductase Standard:** Reconstitute the Human Aldose Reductase Standard (16 ng) with 0.8 ml of **Standard Diluent** to generate a 20 ng/ml standard stock solution. Allow the vial to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting from the standard stock solution (20 ng/ml) 2-fold with equal volume of **EIA Diluent** to produce 10, 5, 2.5, 1.25, 0.625, 0.313, and 0.156 ng/ml solutions. EIA Diluent serves as the zero standard (0 ng/ml). Aliquot remaining stock solution to limit repeated freeze-thaw cycles. This solution should be stored at -20°C and **used within 30 days**.

Standard Point	Dilution	[AR] (ng/ml)
P1	1 part Standard (20 ng/ml) + 1 part EIA Diluent	10.0
P2	1 part P1 + 1 part EIA Diluent	5.0
P3	1 part P2 + 1 part EIA Diluent	2.5
P4	1 part P3 + 1 part EIA Diluent	1.25
P5	1 part P4 + 1 part EIA Diluent	0.625
P6	1 part P5 + 1 part EIA Diluent	0.313
P7	1 part P6 + 1 part EIA Diluent	0.156
P8	EIA Diluent	0.0

- Biotinylated Human Aldose Reductase Antibody (40x):** Spin down the antibody briefly and dilute the desired amount of the antibody 40-fold with EIA Diluent to produce a 1x solution. The undiluted antibody should be stored at -20°C.
- Wash Buffer Concentrate (20x):** Dilute the Wash Buffer Concentrate 20-fold with reagent grade water to produce a 1x solution. When diluting the concentrate, make sure to rinse the bottle thoroughly to extract any precipitates left in the bottle. Mix the 1x solution gently until the crystals have completely dissolved.
- SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 100-fold with EIA Diluent to produce a 1x solution. The undiluted conjugate should be stored at -20°C.

Assay Procedure

- Prepare all reagents, standard solutions, and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-25°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Human Aldose Reductase Standard or sample to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 2 hours. Start the timer after the last addition.
- Wash the microplate manually or automatically using a microplate washer. Invert the plate and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If washing manually, wash five times with 200 µl of Wash Buffer per well. Invert the plate each time and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If using a microplate washer, wash six times with 300 µl of Wash Buffer per well; invert the plate and hit 4-5 times on absorbent material to completely remove the liquid.
- Add 50 µl of Biotinylated Human Aldose Reductase Antibody to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 2 hours.
- Wash the microplate as described above.
- Add 50 µl of SP Conjugate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.
- Add 50 µl of Chromogen Substrate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Incubate in ambient light for 20 minutes or until the optimal blue color density develops.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow. Gently tap plate to ensure thorough mixing. Break any bubbles that may have formed.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance (OD) on the y-axis. The best fit line can be determined by regression analysis using log-log or four-parameter logistic curve fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.


Typical Data

- The typical data is provided for reference only. Individual laboratory means may vary from the values listed. Variations between laboratories may be caused by technique differences.

Standard Point	ng/ml	OD	Average OD
P1	10.0	1.927 1.831	1.879
P2	5.0	1.239 1.155	1.197
P3	2.5	0.680 0.622	0.651
P4	1.25	0.346 0.376	0.361
P5	0.625	0.207 0.185	0.196
P6	0.313	0.128 0.114	0.121
P7	0.156	0.090 0.080	0.085
P8	0.0	0.050 0.046	0.048
Sample: Pooled Normal Sodium Citrate Plasma (1x)		0.331 0.367	0.349
Sample: Pooled Normal Serum (1x)		0.323 0.295	0.309

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



Performance Characteristics

- This assay recognizes both natural and recombinant human aldose reductase.
- The minimum detectable dose of human aldose reductase as calculated by 2SD from the mean of a zero standard was established to be 88 pg/ml.
- Intra-assay precision was determined by testing three plasma samples twenty times in one assay.
- Inter-assay precision was determined by testing three plasma samples in twenty assays.

	Intra-Assay Precision			Inter-Assay Precision		
Sample	1	2	3	1	2	3
n	20	20	20	20	20	20
CV (%)	5.2%	3.9%	4.0%	11.2%	10.0%	9.4%
Average CV (%)	4.4%			10.2%		

Recovery

Standard Added Value	0.6 – 5.0 ng/ml
Recovery %	88 – 112%
Average Recovery %	96%

Linearity

- Plasma and serum samples were serially diluted to test for linearity.

Average Percentage of Expected Value (%)		
Sample Dilution	Plasma	Serum
1x	90%	91%
2x	107%	101%
4x	107%	109%

Cross-Reactivity

Species	Cross-Reactivity (%)
Canine	10%
Bovine	None
Equine	10%
Monkey	70%
Mouse	None
Rat	20%
Swine	25%
Rabbit	None

- No significant cross-reactivity observed with AKR1A1, AKR1B10, AKR1C1, AKR1C3, and AKR1C4.

Troubleshooting

Issue	Causes	Course of Action
Low Precision	Use of improper components	- Check the expiration date listed before use. - Do not interchange components from different lots.
	Improper wash step	- Check that the correct wash buffer is being used. - Check that all wells are empty after aspiration. - Check that the microplate washer is dispensing properly. - If washing by pipette, check for proper pipetting technique.
	Splashing of reagents while loading wells	Pipette properly in a controlled and careful manner.
	Inconsistent volumes loaded into wells	- Pipette properly in a controlled and careful manner. - Check pipette calibration. - Check pipette for proper performance.
	Insufficient mixing of reagent dilutions	- Thoroughly agitate the lyophilized components after reconstitution. - Thoroughly mix dilutions.
	Improperly sealed microplate	- Check the microplate pouch for proper sealing. - Check that the microplate pouch has no punctures. - Check that three desiccants are inside the microplate pouch prior to sealing.

Unexpectedly Low or High Signal Intensity	Microplate was left unattended between steps	Each step of the procedure should be performed uninterrupted.
	Omission of step	Consult the provided procedure for complete list of steps.
	Steps performed in incorrect order	Consult the provided procedure for the correct order.
	Insufficient amount of reagents added to wells	- Check pipette calibration. - Check pipette for proper performance.
	Wash step was skipped	Consult the provided procedure for all wash steps.
	Improper wash buffer	Check that the correct wash buffer is being used.
	Improper reagent preparation	Consult reagent preparation section for the correct dilutions of all reagents.
Deficient Standard Curve Fit	Insufficient or prolonged incubation periods	Consult the provided procedure for correct incubation time.
	Non-optimal sample dilution	- Sandwich ELISA: If samples generate OD values higher than the highest standard point (P1), dilute samples further and repeat the assay. - Competitive ELISA: If samples generate OD values lower than the highest standard point (P1), dilute samples further and repeat the assay. - User should determine the optimal dilution factor for samples.
	Contamination of reagents	A new tip must be used for each addition of different samples or reagents during the assay procedure.
	Contents of wells evaporate	Verify that the sealing film is firmly in place before placing the assay in the incubator or at room temperature.
	Improper pipetting	- Pipette properly in a controlled and careful manner. - Check pipette calibration. - Check pipette for proper performance.
	Insufficient mixing of reagent dilutions	- Thoroughly agitate the lyophilized components after reconstitution. - Thoroughly mix dilutions.

References

- (1) Graham A *et al.* (1991) *J Biol Chem.* 266(11):6872-7.
- (2) Ramana KV *et al.* (2004) *FEBS Lett.* 570(1-3):189-94.

Version 1.3R1